Posted by Zbych on Sun, 13 Mar 2022 19:44:37 GMT
View Forum Message <> Reply to Message

First step should be checking if somebody already created communication library.
If not (which I doubt), check HID usage tables reported by device
https://www.usb.org/hid
  https://www.usb.org/sites/default/files/documents/hut1_12v2. pdf

If your device reports back as "generic hid", you need to get a description of "reports" that device
recognizes, since they are device specific.

To be honest, I really doubt that your oscilloscope uses USB HID class, it is just to slow to send
measurement information to a PC.